



Idaho State Police Forensic Services

CERTIFICATE OF ANALYSIS

The Department of Law Enforcement Idaho State Police Forensic Services hereby certifies that Simulator Solution Lot Number 0000007109 to be used to conduct calibration checks within the State of Idaho in accordance with the policies and/or procedures promulgated by the Department governing breath alcohol examinations. This lot has a target value of 0.206 with a range of 0.186 to 0.226 grams of ethyl alcohol/210 liters of vapor.

The expiration date for this lot number is April 3, 2009 at 11:59 PM.

December 19, 2007
Date

David A. Laycock
Forensic Scientist

STATE OF IDAHO)
) ss.
County of Ada)

On this 19th of December, in the year 2007, before me, Jane Davenport, a notary public, personally appeared David A. Laycock, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledge to me that he executed the same as such Scientist.

Jane Davenport Notary Public

My Commission Expires: 5 Jan 13

